



Instructions for Use

Integrity™ Bone Staple Fixation System

This Instructions for Use is intended exclusively for distribution within the United States.

In an effort to best meet the needs of our customers and minimize waste, these Instructions for Use are being provided in electronic format. This document is subject to change; the most current version of this IFU is available online. If unsure if using the latest revision, please reprint the IFU at www.anikaifu.com. If a paper copy is preferred it may be requested, free of charge, by contacting Anika Therapeutics, Inc. at www.anikaifu.com. The onus resides with the user to ensure that the most up to date IFU is used.

Products may not be available in all markets. Product availability is subject to regulatory clearance in individual markets. Please contact your local representative if you have questions about product availability in your area.



Caution: Federal Law restricts this device to sale by or on the order of a physician.

Read these instructions completely prior to use.

DEVICE DESCRIPTION

The Integrity™ Bone Staple is a staple-shaped tack with barbed ends and is composed of polyether ether ketone (PEEK) material. The Integrity™ Bone Staple is used in conjunction with the associated Delivery Instruments and provides fixation of soft tissue grafts or reinforcement meshes to bone. The Integrity™ Bone Staples are provided sterile for single-use only and are packaged in a caddy for placement and presentation.

The Integrity Bone Staple Fixation System includes the Integrity Implant Fixation Kit, including PEEK Bone Staples and PLGA Tissue Tacks, and the Integrity Retractable Implant Insertor.

HOW SUPPLIED

The Integrity™ Bone Staples are provided sterile for single-use only and are packaged in a caddy within a dual sterile seal packaging configuration. The caddy may also contain other implants, including Tissue Tacks. The Delivery Instruments are packaged separately and provided sterile for single use only. Contents of each package are sterile unless the package is opened or damaged. The Integrity™ Bone Staple and Delivery Instruments, and packaging do not contain natural rubber latex. DO NOT use if package is damaged or if labeling is incomplete or illegible.

INTENDED USE

The Integrity™ Bone Staple Fixation System is intended for the fixation of soft tissue grafts or reinforcement meshes to bone during rotator cuff repairs.

INDICATIONS FOR USE

The Integrity™ Bone Staple Fixation System is indicated for the fixation of soft tissue grafts or reinforcement meshes to bone during rotator cuff repairs.

CONTRAINDICATIONS

- The Integrity™ Bone Staple and Delivery instruments are not designed, sold, or intended for use except as described in the indications for use and is contraindicated in the following situation:
 - The Integrity™ Bone Staples are not indicated to reinforce the strength of any tendon repair.
 - The Integrity™ Bone Staples are not indicated where there is inadequate quality of bone.

WARNINGS

- Do not use if the package is damaged. Do not use if the product sterilization barrier or its packaging is compromised.
- Contents are sterile unless package is opened or damaged. DO NOT RESTERILIZE.
- For single use only.
- Discard any open, unused product. Do not use after the expiration date.
- The surgeon shall be thoroughly familiar with the implants, Delivery Instruments, and surgical technique prior to performing surgery.
- Read these instructions completely prior to use.
- The Integrity™ Delivery instruments are not indicated for use with implants manufactured by any company other than Anika.
- Ensure that all fixation implants are properly secured prior to patient closure.
- Do not apply excessive forces or torque to the instruments. Avoid hard, heavy hammer blows.

MRI SAFETY INFORMATION

Integrity™ Bone Staple is MRI Safe.

STORAGE

- Store at room temperature: 59°F (15°C) to 86°F (30°C). Keep dry.

For caddy packaging (Integrity Implant Fixation Kit, including PEEK Bone Staples and PLGA Tissue Tacks), avoid temperatures more than 104°F (40°C).

- Outer caddy package includes a temperature indicator. Do not use if central circle of the indicator on the product appears black. In the event of a temperature breach the indicator window will turn from white to black.

PRECAUTIONS

- Prior to use, inspect the device to ensure it is not damaged. Do not use a damaged device.
- Hazards associated with reuse include, but are not limited to, patient infection and/or device malfunction. Surgeon or Physician should discuss general risks and potential complications associated with this and any surgical procedure with the patient prior to patient consent.
- Overlapping fixation devices may result in damage to the devices.
- Integrity™ Bone Staples should be placed approximately 4 mm from edge of soft tissue graft or reinforcement mesh to avoid tearing.
- Tips of the Integrity™ Retractable Implant Inserters are sharp, use caution when handling device.
- Insertion of the Integrity™ Bone Staples through excessive tissue or augment thickness may not provide adequate fixation.
- Application of the Integrity™ Bone Staple does not modify the postoperative treatment. The surgeon must determine motion and strength requirements according to standard practice.

POTENTIAL ADVERSE EVENTS

The following are potential adverse events that may occur from the surgical procedure or complications with the Integrity™ Bone Staple or associated Delivery Instrument:

- Allergic Reaction
- Infection
- Pain
- Device may not function as intended.

DIRECTIONS FOR USE

1. Rotate Caddy Cover counterclockwise to reveal fixation implants. Insert the Integrity™ Retractable Implant Inserter into a cavity labeled "STAPLE" aligning trocar needle tips with staple lumens, to retrieve a bone staple. Apply downward pressure on the back of the handle rather than the shaft to ensure tines do not retract due to resistance with the staple.
2. Pass bone staple through the soft tissue graft or reinforcement mesh extracorporeally.
3. Pass the distal end of the delivery instrument through the surgical access Integrity™ Cannula and into position or through a commercially available non-rigid cannula using the Deployment Skid Cannula Access Tool.
4. If using the Integrity™ Cannula, advance the seal down to engage the Integrity™ Cannula and rotate clockwise to secure into position. Locate the bone staple at the desired fixation location and impact proximal handle to seat bone staple into bone. Avoid excessive impaction during insertion as this could lead to instrument and/or implant damage. If insertion resistance is encountered, do not impact harder. Replace the implant, reassess position, and repeat the process.
5. Secure the clear Integrity™ Retractable Implant Inserter ring while pulling the handle to retract / eject the tines from the bone staple implant. Once the tines are fully retracted, use the distal end to assist with placement of the soft tissue graft or reinforcement mesh. Use additional suture fixation techniques and/or devices to finalize placement and secure fixation.
6. Remove the delivery instrument. Confirm adequate fixation has been achieved.
7. Dispose of the devices according to the appropriate environmental health safety guidelines.
8. Refer to the surgical technique guide for additional information.

	CAUTION: U.S. Federal law restricts this device to sale by or on the order of a physician		MR Safe
	Double sterile barrier system		Catalog Number
	Single sterile barrier system		Lot number
	Sterilized using irradiation		Use by date
	Do not resterilize		Manufacturer
	Do not reuse		Date of Manufacture
	Do not use if package is damaged and consult instructions for use		Unique Device Identifier
	Upper limit of temperature	 www.anikaifu.com	Consult electronic instructions for use
	Keep Dry		

FOR FURTHER INFORMATION

If further information on this product is needed, please contact Anika Customer Service at 1-888-721-1600 in the U.S., or an authorized representative.



Anika Therapeutics, Inc.
32 Wiggins Avenue
Bedford, MA 01730, USA
Made in USA

AML-300-908 REV-A 2026-09